
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002195
11:43 01/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 1:57
STERILIZING 2:54
136.4 C 223kPa 2:54
136.4 C 226kPa 3:24
136.3 C 221kPa 3:54
136.1 C 222kPa 4:24
136.3 C 224kPa 4:54
136.4 C 226kPa 5:24
136.2 C 224kPa 5:54
136.1 C 223kPa 6:24
136.1 C 223kPa 6:24
VENTING 6:25
AIR DRYING 7:17
CYCLE COMPLETE 21:42

Digital Signature #
2A9C16D00C3CAD27
