
Midmark M11 - v1.0.5
Total Cycles: 105

Name: RITTER2
Sterilizer ID: V2313228
06/16/2021 08:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.0 F 0.0
2:00 75.0 F 0.2
4:00 89.9 F 1.1
6:00 106.1 F 3.1
8:00 136.4 F 5.6
10:00 164.3 F 8.6
12:00 187.1 F 11.8
14:00 201.3 F 3.2
16:00 210.3 F 6.6
18:00 220.7 F 10.0
20:00 232.6 F 11.2
22:00 232.9 F 9.9
24:00 238.1 F 10.3
26:00 244.0 F 13.0
28:00 254.3 F 18.4
30:00 263.7 F 24.1
32:00 272.9 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.5
1:00 272.4 F 29.7
2:00 272.6 F 30.1
3:00 271.6 F 29.5
4:00 272.2 F 29.8

Min 271.5 F 29.4
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.6 F 0.0
5:00 201.2 F 0.0
10:00 199.5 F 0.0
15:00 196.3 F 0.0
20:00 195.1 F 0.0
25:00 194.3 F 0.0
30:00 194.1 F 0.0

Drying Complete

FILLING: 1:02
HEATING: 33:08
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:11:34
CYCLE COMPLETE

06/16/2021 09:25 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____