
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000126
8:00 08/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:12
PRESSURIZING 4:13
STERILIZING 5:46
136.0 C 220kPa 5:46
136.2 C 223kPa 6:16
136.2 C 224kPa 6:46
136.1 C 224kPa 7:16
136.3 C 225kPa 7:46
136.3 C 225kPa 8:16
136.2 C 224kPa 8:46
136.2 C 224kPa 9:16
136.2 C 224kPa 9:16
VENTING 9:16
AIR DRYING 10:23
CYCLE COMPLETE 13:19

Digital Signature #
60E2FE2BC67E63CC
