
Midmark M11 - v1.0.0
Total Cycles: 3636

Name: RITTER 1
Sterilizer ID: V2178198
08/10/2023 01:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.1 F 0.0
2:00 190.4 F 4.4
4:00 213.4 F 10.2
6:00 224.6 F 6.3
8:00 236.9 F 12.3
10:00 238.3 F 10.4
12:00 236.9 F 9.3
14:00 252.3 F 16.7
16:00 264.9 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.4 F 29.0
2:00 271.5 F 28.6
3:00 272.5 F 29.2
4:00 271.7 F 28.5

Min 271.4 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 225.0 F 0.0
10:00 239.5 F 0.0
15:00 234.3 F 0.0
20:00 225.7 F 0.0
25:00 222.3 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 2:31
HEATING: 17:46
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:57:37
CYCLE COMPLETE

08/10/2023 02:06 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____