
Midmark M11 - v1.0.0
Total Cycles: 5002

Name: RITTER 1
Sterilizer ID: V2178198
10/16/2024 04:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.3 F 0.0
2:00 170.1 F 0.7
4:00 190.6 F 4.5
6:00 209.6 F 9.2
8:00 221.8 F 10.0
10:00 229.1 F 9.1
12:00 235.4 F 10.3
14:00 241.2 F 11.9
16:00 240.2 F 10.3
18:00 253.2 F 16.8
20:00 263.8 F 23.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.3
1:00 272.2 F 28.9
2:00 272.8 F 28.8
3:00 271.3 F 28.2
4:00 272.3 F 29.1

Min 271.2 F 28.2
Max 273.1 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.5 F 0.0
5:00 228.3 F 0.0
10:00 247.2 F 0.0
15:00 247.4 F 0.0
20:00 235.9 F 0.0
25:00 231.9 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 21:57
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:01:02
CYCLE COMPLETE

10/16/2024 05:28 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____