
Midmark M11 - v1.0.0
Total Cycles: 4732

Name: RITTER 1
Sterilizer ID: V2178198
07/16/2024 08:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 94.6 F 0.0
4:00 134.1 F 0.3
6:00 154.2 F 2.2
8:00 153.0 F 4.6
10:00 212.1 F 7.6
12:00 211.2 F 11.7
14:00 213.0 F 3.1
16:00 225.3 F 6.4
18:00 231.5 F 9.9
20:00 234.3 F 11.0
22:00 236.6 F 9.8
24:00 238.4 F 9.5
26:00 244.9 F 12.7
28:00 255.9 F 18.6
30:00 266.2 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.9 F 29.0
2:00 271.2 F 28.3
3:00 272.5 F 29.1
4:00 271.8 F 28.4

Min 271.2 F 28.3
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.0 F 0.0
5:00 218.9 F 0.0
10:00 210.7 F 0.0
15:00 242.5 F 0.0
20:00 235.1 F 0.0
25:00 230.1 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 31:47
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:10:01
CYCLE COMPLETE

07/16/2024 09:33 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____