
Midmark M11 - v1.0.0
Total Cycles: 2786

Name: RITTER 1
Sterilizer ID: V2178198
01/17/2023 02:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 214.0 F 0.0
2:00 208.2 F 3.8
4:00 215.1 F 8.8
6:00 224.0 F 7.8
8:00 232.2 F 9.7
10:00 238.6 F 8.8
12:00 244.7 F 13.5
14:00 244.1 F 12.5
16:00 257.2 F 19.4
18:00 268.9 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.4 F 29.1
2:00 271.3 F 28.5
3:00 272.4 F 29.3
4:00 271.8 F 28.7

Min 271.2 F 28.4
Max 272.4 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.5 F 0.0
5:00 234.5 F 0.0
10:00 232.7 F 0.0
15:00 228.5 F 0.0
20:00 227.7 F 0.0
25:00 229.6 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 19:05
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:58
CYCLE COMPLETE

01/17/2023 03:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____