

-----  
Midmark M11 - v1.0.0  
Total Cycles: 757

Name: RITTER 1  
Sterilizer ID: V2178198  
03/31/2021 06:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.7 F 0.0  
2:00 71.8 F 0.0  
4:00 75.5 F 0.1  
6:00 98.7 F 0.1  
8:00 141.2 F 0.2  
10:00 179.4 F 0.4  
12:00 206.3 F 3.5  
14:00 220.9 F 6.7  
16:00 227.9 F 10.0  
18:00 235.7 F 10.6  
20:00 237.6 F 9.8  
22:00 237.9 F 9.2  
24:00 245.2 F 13.0  
26:00 243.3 F 11.7  
28:00 256.4 F 18.5  
30:00 267.7 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.0  
1:00 272.3 F 28.3  
2:00 272.6 F 28.8  
3:00 271.4 F 28.0  
4:00 272.5 F 28.5

Min 271.4 F 27.8  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 171.6 F 0.0  
5:00 240.2 F 0.0  
10:00 228.7 F 0.0  
15:00 225.5 F 0.0  
20:00 228.3 F 0.0  
25:00 191.7 F 0.0  
30:00 214.1 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 31:18  
STERILIZING: 4:00  
VENTING: 2:59  
DRYING: 30:00  
TOTAL CYCLE: 01:09:06  
CYCLE COMPLETE

03/31/2021 07:35 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01