
Midmark M11 - v1.0.0
Total Cycles: 3687

Name: RITTER 1
Sterilizer ID: V2178198
08/28/2023 08:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.6 F 0.0
2:00 88.1 F 0.0
4:00 83.2 F 0.0
6:00 121.5 F 1.3
8:00 141.8 F 3.6
10:00 161.6 F 6.6
12:00 196.9 F 10.1
14:00 212.0 F 8.9
16:00 220.3 F 5.2
18:00 226.4 F 8.5
20:00 230.9 F 12.2
22:00 231.6 F 7.5
24:00 242.4 F 12.3
26:00 238.7 F 9.7
28:00 250.9 F 15.5
30:00 261.2 F 21.4
32:00 270.9 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.4 F 28.6
2:00 271.5 F 28.4
3:00 272.3 F 28.4
4:00 271.6 F 28.6

Min 271.1 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.3 F 0.0
5:00 217.8 F 0.0
10:00 218.9 F 0.0
15:00 224.9 F 0.0
20:00 215.8 F 0.0
25:00 223.7 F 0.0
30:00 225.7 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 32:46
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:10:52
CYCLE COMPLETE

08/28/2023 09:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____