

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 006681  
11:48 04/12/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:33  
PRESSURIZING 2:35  
STERILIZING 3:38  
136.5 C 220kPa 3:38  
136.3 C 224kPa 4:08  
136.4 C 221kPa 4:38  
136.4 C 223kPa 5:08  
136.4 C 223kPa 5:38  
136.2 C 223kPa 6:08  
136.4 C 223kPa 6:38  
136.3 C 224kPa 7:08  
136.3 C 224kPa 7:08  
VENTING 7:08  
AIR DRYING 7:58  
CYCLE COMPLETE 8:10

-----  
Digital Signature #  
5DF4D949C90890F3  
-----