
Midmark M11 - v1.0.5
Total Cycles: 3272

Name: RITTER2
Sterilizer ID: V2313228
01/03/2024 03:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.1 F 0.0
2:00 138.0 F 1.1
4:00 160.5 F 4.0
6:00 182.6 F 7.8
8:00 199.6 F 12.1
10:00 224.4 F 5.9
12:00 235.1 F 10.5
14:00 232.5 F 8.1
16:00 242.1 F 13.1
18:00 242.5 F 12.0
20:00 254.9 F 18.3
22:00 266.0 F 25.0
24:00 272.5 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.4
1:00 272.4 F 29.5
2:00 272.4 F 29.9
3:00 271.6 F 29.7
4:00 272.2 F 29.5

Min 271.3 F 29.3
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.1 F 0.0
5:00 238.5 F 0.0
10:00 193.3 F 0.0
15:00 185.8 F 0.0
20:00 182.7 F 0.0
25:00 180.3 F 0.0
30:00 178.9 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 24:25
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:03:09
CYCLE COMPLETE

01/03/2024 04:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____