
Midmark M11 - v1.0.0
Total Cycles: 1187

Name: RITTER 1
Sterilizer ID: V2178198
09/20/2021 10:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.9 F 0.0
2:00 197.4 F 2.7
4:00 200.4 F 7.6
6:00 217.4 F 12.7
8:00 225.7 F 7.2
10:00 238.4 F 12.3
12:00 237.8 F 9.8
14:00 239.1 F 8.9
16:00 249.7 F 15.0
18:00 261.5 F 21.6
20:00 272.2 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.7 F 28.8
2:00 271.6 F 28.1
3:00 272.8 F 28.9
4:00 272.0 F 28.3

Min 271.6 F 28.0
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.8 F 0.0
5:00 227.3 F 0.0
10:00 202.4 F 0.0
15:00 202.1 F 0.0
20:00 199.2 F 0.0
25:00 208.8 F 0.0
30:00 219.5 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 20:29
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:52
CYCLE COMPLETE

09/20/2021 11:22 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____