
Midmark M11 - v1.0.5
Total Cycles: 529

Name: RITTER2
Sterilizer ID: V2313228
11/09/2021 01:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.0 F 0.0
2:00 181.5 F 3.7
4:00 187.8 F 8.3
6:00 206.6 F 12.6
8:00 222.9 F 7.9
10:00 235.5 F 13.3
12:00 238.2 F 11.0
14:00 237.6 F 9.8
16:00 251.3 F 16.5
18:00 262.8 F 23.3
20:00 272.5 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.7 F 29.8
2:00 271.5 F 29.0
3:00 272.1 F 29.9
4:00 272.4 F 29.6

Min 271.4 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.4 F 0.0
5:00 225.2 F 0.0
10:00 237.1 F 0.0
15:00 231.0 F 0.0
20:00 224.1 F 0.0
25:00 222.1 F 0.0
30:00 221.1 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 20:14
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:21
CYCLE COMPLETE

11/09/2021 02:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____