
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002191
9:23 01/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:59
PRESSURIZING 2:05
STERILIZING 3:08
136.0 C 220kPa 3:08
135.9 C 222kPa 3:38
136.2 C 224kPa 4:08
136.4 C 225kPa 4:38
136.4 C 223kPa 5:08
136.3 C 224kPa 5:38
136.2 C 224kPa 6:08
136.4 C 224kPa 6:38
136.4 C 224kPa 6:38
VENTING 6:38
AIR DRYING 7:28
CYCLE COMPLETE 27:55

Digital Signature #
BBF185D03DD6BEE2
