
Midmark M11 - v1.0.5
Total Cycles: 3660

Name: RITTER2
Sterilizer ID: V2313228
05/30/2024 01:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 86.2 F 0.0
2:00 87.2 F 0.0
4:00 108.4 F 0.8
6:00 133.4 F 2.9
8:00 150.4 F 5.6
10:00 169.5 F 8.8
12:00 186.4 F 12.1
14:00 209.0 F 5.0
16:00 220.5 F 8.8
18:00 231.4 F 12.9
20:00 232.1 F 9.2
22:00 238.5 F 10.9
24:00 242.7 F 12.4
26:00 253.4 F 18.1
28:00 263.6 F 24.0
30:00 272.8 F 30.3

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.7
1:00 271.8 F 30.1
2:00 272.5 F 30.0
3:00 271.3 F 29.6
4:00 272.5 F 30.4

Min 271.3 F 29.5
Max 273.0 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.8 F 0.0
5:00 200.3 F 0.0
10:00 205.3 F 0.0
15:00 204.2 F 0.0
20:00 200.2 F 0.0
25:00 201.3 F 0.0
30:00 201.6 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 30:15
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:08:47
CYCLE COMPLETE

05/30/2024 02:30 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____