
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 003022
8:46 26/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:07
STERILIZING 4:09
136.4 C 224kPa 4:09
136.3 C 223kPa 4:39
136.4 C 222kPa 5:09
136.2 C 223kPa 5:39
136.3 C 220kPa 6:09
136.1 C 221kPa 6:39
136.4 C 223kPa 7:09
136.4 C 221kPa 7:39
136.4 C 221kPa 7:39
VENTING 7:39
AIR DRYING 8:33
CYCLE COMPLETE 8:46

Digital Signature #
BC0359970FBA9958
