

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3689

Name: RITTER 1  
Sterilizer ID: V2178198  
08/28/2023 10:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.9 F 0.0  
2:00 185.9 F 3.1  
4:00 182.8 F 7.0  
6:00 208.6 F 12.0  
8:00 219.1 F 6.2  
10:00 234.0 F 11.2  
12:00 233.1 F 7.9  
14:00 243.9 F 12.4  
16:00 244.0 F 12.3  
18:00 256.3 F 18.7  
20:00 267.1 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.3  
1:00 272.6 F 29.1  
2:00 271.6 F 28.2  
3:00 272.7 F 28.9  
4:00 271.5 F 28.3  
  
Min 271.5 F 28.2  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.3 F 0.0  
5:00 213.6 F 0.0  
10:00 222.5 F 0.0  
15:00 230.8 F 0.0  
20:00 221.9 F 0.0  
25:00 218.5 F 0.0  
30:00 217.7 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 21:20  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:00:07  
CYCLE COMPLETE

08/28/2023 11:44 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_