
Midmark M11 - v1.0.0
Total Cycles: 2478

Name: RITTER 1
Sterilizer ID: V2178198
10/24/2022 12:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.9 F 0.0
2:00 159.4 F 3.2
4:00 171.1 F 7.1
6:00 192.9 F 11.7
8:00 213.0 F 5.0
10:00 224.8 F 9.8
12:00 231.9 F 8.6
14:00 239.4 F 11.3
16:00 236.9 F 9.1
18:00 250.4 F 15.5
20:00 261.6 F 21.9
22:00 272.0 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.1
2:00 271.7 F 28.4
3:00 271.8 F 28.9
4:00 272.2 F 29.0

Min 271.2 F 28.3
Max 272.4 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.5 F 0.0
5:00 228.4 F 0.0
10:00 236.1 F 0.0
15:00 230.2 F 0.0
20:00 228.4 F 0.0
25:00 229.9 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:35
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:55
CYCLE COMPLETE

10/24/2022 01:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____