
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001970
7:28 22/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:08
STERILIZING 4:15
136.2 C 221kPa 4:15
136.2 C 223kPa 4:45
136.4 C 225kPa 5:15
136.4 C 224kPa 5:45
136.4 C 224kPa 6:15
136.4 C 222kPa 6:45
136.4 C 224kPa 7:15
136.4 C 224kPa 7:45
136.4 C 224kPa 7:45
VENTING 7:45
AIR DRYING 8:36
CYCLE COMPLETE 38:36

Digital Signature #
773A6E46C7B502F0
