

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002793  
10:07 08/12/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:14  
PRESSURIZING 3:17  
STERILIZING 4:22  
136.5 C 221kPa 4:22  
136.2 C 222kPa 4:52  
136.4 C 222kPa 5:22  
136.4 C 224kPa 5:52  
136.4 C 223kPa 6:22  
136.4 C 222kPa 6:52  
136.0 C 220kPa 7:22  
136.4 C 224kPa 7:52  
136.4 C 224kPa 7:52  
VENTING 7:52  
AIR DRYING 8:43  
CYCLE COMPLETE 8:54

-----  
Digital Signature #  
262C76D633C1199D  
-----