
Midmark M11 - v1.0.0
Total Cycles: 5399

Name: RITTER 1
Sterilizer ID: V2178198
03/07/2025 01:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.9 F 0.0
2:00 136.6 F 1.8
4:00 165.7 F 4.9
6:00 175.3 F 8.5
8:00 195.0 F 12.6
10:00 218.6 F 4.7
12:00 230.9 F 8.6
14:00 236.8 F 13.0
16:00 235.8 F 9.1
18:00 240.1 F 10.5
20:00 244.2 F 12.6
22:00 255.7 F 18.6
24:00 266.4 F 25.0
26:00 272.6 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.7 F 29.1
2:00 272.4 F 29.5
3:00 271.7 F 28.9
4:00 272.8 F 29.3

Min 271.7 F 28.8
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.7 F 0.0
5:00 204.5 F 0.0
10:00 215.6 F 0.0
15:00 195.0 F 0.0
20:00 243.8 F 0.0
25:00 233.6 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 26:25
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:05:03
CYCLE COMPLETE

03/07/2025 02:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____