
Midmark M11 - v1.0.5
Total Cycles: 2564

Name: RITTER2
Sterilizer ID: V2313228
06/26/2023 09:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.8 F 0.0
2:00 163.5 F 0.0
4:00 166.1 F 2.2
6:00 191.6 F 6.7
8:00 213.5 F 11.6
10:00 222.6 F 7.0
12:00 235.6 F 12.3
14:00 236.1 F 10.0
16:00 236.6 F 8.8
18:00 249.3 F 15.4
20:00 261.1 F 22.2
22:00 271.8 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.6 F 29.7
2:00 271.5 F 29.0
3:00 272.7 F 29.8
4:00 272.0 F 29.2

Min 271.5 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.9 F 0.0
5:00 230.1 F 0.0
10:00 238.4 F 0.0
15:00 229.7 F 0.0
20:00 225.7 F 0.0
25:00 229.6 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 22:34
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:48
CYCLE COMPLETE

06/26/2023 10:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____