
Midmark M11 - v1.0.0
Total Cycles: 3530

Name: RITTER 1
Sterilizer ID: V2178198
07/14/2023 04:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 85.1 F 0.0
2:00 99.3 F 0.0
4:00 119.2 F 0.5
6:00 146.3 F 1.6
8:00 167.4 F 3.5
10:00 189.1 F 6.1
12:00 199.2 F 9.4
14:00 210.0 F 13.1
16:00 220.3 F 5.8
18:00 230.3 F 9.6
20:00 239.1 F 12.4
22:00 238.9 F 10.5
24:00 236.1 F 8.4
26:00 248.0 F 14.1
28:00 258.8 F 20.0
30:00 268.6 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.4 F 28.8
2:00 271.4 F 28.1
3:00 272.3 F 29.0
4:00 271.8 F 28.4

Min 271.2 F 28.1
Max 272.4 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.1 F 0.0
5:00 206.3 F 0.0
10:00 203.2 F 0.0
15:00 192.6 F 0.0
20:00 195.9 F 0.0
25:00 195.7 F 0.0
30:00 198.9 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 31:15
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:09:33
CYCLE COMPLETE

07/14/2023 05:24 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____