
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002467
15:55 28/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:35
STERILIZING 3:37
136.0 C 222kPa 3:37
136.4 C 225kPa 4:07
136.4 C 226kPa 4:37
136.4 C 226kPa 5:07
136.4 C 225kPa 5:37
136.2 C 224kPa 6:07
136.4 C 226kPa 6:37
136.4 C 224kPa 7:07
136.4 C 224kPa 7:07
VENTING 7:07
AIR DRYING 8:02
CYCLE COMPLETE 17:34

Digital Signature #
DEC946C17B925876
