
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002288
8:51 19/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:54
STERILIZING 4:03
136.4 C 223kPa 4:03
136.4 C 224kPa 4:33
136.4 C 224kPa 5:03
136.0 C 222kPa 5:33
136.1 C 222kPa 6:03
136.3 C 223kPa 6:33
136.2 C 222kPa 7:03
136.3 C 222kPa 7:33
136.3 C 222kPa 7:33
VENTING 7:33
AIR DRYING 8:24
CYCLE COMPLETE 38:24

Digital Signature #
3DEF4D7FBC99061A
