
Midmark M11 - v1.0.0
Total Cycles: 2794

Name: RITTER 1
Sterilizer ID: V2178198
01/19/2023 08:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.0 F 0.0
2:00 79.1 F 0.0
4:00 94.2 F 0.0
6:00 129.0 F 0.0
8:00 158.6 F 1.6
10:00 182.4 F 3.8
12:00 195.6 F 6.6
14:00 211.7 F 9.8
16:00 224.6 F 13.5
18:00 225.9 F 7.0
20:00 236.0 F 11.1
22:00 232.2 F 7.4
24:00 245.7 F 13.5
26:00 243.7 F 12.0
28:00 256.2 F 18.4
30:00 267.0 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.2 F 28.5
2:00 271.9 F 28.6
3:00 271.9 F 28.3
4:00 272.2 F 28.9

Min 271.4 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 171.2 F 0.0
5:00 232.5 F 0.0
10:00 226.6 F 0.0
15:00 233.1 F 0.0
20:00 227.5 F 0.0
25:00 230.1 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:27
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:09:22
CYCLE COMPLETE

01/19/2023 09:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____