
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000248
14:37 29/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:56
PRESSURIZING 3:57
STERILIZING 5:26
136.4 C 224kPa 5:26
136.4 C 233kPa 5:56
136.4 C 233kPa 6:26
136.4 C 232kPa 6:56
136.4 C 233kPa 7:26
136.4 C 233kPa 7:56
136.4 C 233kPa 8:26
136.4 C 232kPa 8:56
136.4 C 232kPa 8:56
VENTING 8:56
AIR DRYING 10:00
CYCLE COMPLETE 11:52

Digital Signature #
E55B876F7D38E388
