

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2843

Name: RITTER 1  
Sterilizer ID: V2178198  
02/01/2023 02:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.3 F 0.0  
2:00 193.1 F 4.5  
4:00 204.7 F 9.7  
6:00 217.8 F 6.0  
8:00 232.2 F 10.3  
10:00 232.1 F 7.4  
12:00 244.9 F 13.0  
14:00 244.4 F 12.4  
16:00 257.2 F 19.3  
18:00 268.7 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.6 F 29.1  
2:00 271.7 F 28.3  
3:00 272.3 F 29.2  
4:00 272.4 F 28.9  
  
Min 271.5 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.3 F 0.0  
5:00 226.8 F 0.0  
10:00 236.2 F 0.0  
15:00 232.2 F 0.0  
20:00 225.3 F 0.0  
25:00 225.9 F 0.0  
30:00 225.7 F 0.0  
Drying Complete

FILLING: 2:01  
HEATING: 19:11  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:26  
CYCLE COMPLETE

02/01/2023 03:41 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_