
Midmark M11 - v1.0.5
Total Cycles: 1312

Name: RITTER2
Sterilizer ID: V2313228
07/18/2022 11:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.0 F 0.0
2:00 170.4 F 0.1
4:00 192.0 F 2.7
6:00 215.8 F 7.6
8:00 231.6 F 12.7
10:00 232.4 F 9.0
12:00 238.5 F 10.3
14:00 243.9 F 13.1
16:00 243.8 F 12.5
18:00 256.8 F 19.4
20:00 268.4 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.7 F 29.5
2:00 271.5 F 28.7
3:00 272.1 F 29.6
4:00 272.4 F 29.3

Min 271.2 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.0 F 0.0
5:00 238.4 F 0.0
10:00 229.0 F 0.0
15:00 223.8 F 0.0
20:00 213.7 F 0.0
25:00 230.0 F 0.0
30:00 214.3 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 21:12
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:22
CYCLE COMPLETE

07/18/2022 12:26 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____