
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002035
11:02 13/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:34
STERILIZING 3:38
136.4 C 225kPa 3:38
136.4 C 226kPa 4:08
136.4 C 223kPa 4:38
136.4 C 225kPa 5:08
136.1 C 222kPa 5:38
136.2 C 224kPa 6:08
136.1 C 223kPa 6:38
136.4 C 225kPa 7:08
136.4 C 225kPa 7:08
VENTING 7:08
AIR DRYING 7:59
CYCLE COMPLETE 14:59

Digital Signature #
FB1B96CE39FB1672
