
Midmark M11 - v1.0.0
Total Cycles: 5338

Name: RITTER 1
Sterilizer ID: V2178198
02/13/2025 09:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.6 F 0.0
2:00 74.7 F 0.4
4:00 93.1 F 2.0
6:00 127.7 F 4.4
8:00 153.0 F 7.2
10:00 172.8 F 10.5
12:00 182.4 F 5.6
14:00 204.6 F 3.8
16:00 216.3 F 7.0
18:00 227.5 F 10.8
20:00 227.5 F 6.9
22:00 237.7 F 10.6
24:00 234.5 F 7.8
26:00 247.3 F 14.1
28:00 258.6 F 20.2
30:00 268.9 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.9 F 29.2
2:00 271.9 F 28.3
3:00 272.7 F 29.3
4:00 272.5 F 28.8

Min 271.6 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 166.4 F 0.0
5:00 190.2 F 0.0
10:00 197.2 F 0.0
15:00 202.2 F 0.0
20:00 201.2 F 0.0
25:00 193.1 F 0.0
30:00 193.3 F 0.0
Drying Complete

FILLING: 0:41
HEATING: 31:11
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:09:20
CYCLE COMPLETE

02/13/2025 10:29 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____