
Midmark M11 - v1.0.0
Total Cycles: 4307

Name: RITTER 1
Sterilizer ID: V2178198
02/16/2024 08:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.3 F 0.0
2:00 75.9 F 0.2
4:00 80.7 F 1.0
6:00 113.2 F 3.0
8:00 147.8 F 5.7
10:00 173.8 F 8.8
12:00 194.8 F 12.2
14:00 209.9 F 2.8
16:00 218.3 F 5.9
18:00 226.4 F 9.1
20:00 235.5 F 12.9
22:00 233.7 F 8.2
24:00 243.5 F 12.8
26:00 239.7 F 10.3
28:00 251.4 F 15.9
30:00 261.4 F 21.6
32:00 270.9 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.1 F 28.5
2:00 272.6 F 28.8
3:00 272.7 F 29.1
4:00 271.9 F 29.0

Min 271.5 F 28.5
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.6 F 0.0
5:00 235.6 F 0.0
10:00 245.5 F 0.0
15:00 217.9 F 0.0
20:00 196.3 F 0.0
25:00 206.9 F 0.0
30:00 212.2 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 33:25
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:11:41
CYCLE COMPLETE

02/16/2024 10:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____