
Midmark M11 - v1.0.0
Total Cycles: 1319

Name: RITTER 1
Sterilizer ID: V2178198
10/29/2021 08:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.7 F 0.0
2:00 153.4 F 1.1
4:00 181.8 F 3.4
6:00 193.6 F 8.2
8:00 207.7 F 13.4
10:00 220.1 F 6.4
12:00 232.8 F 11.0
14:00 233.6 F 7.3
16:00 244.6 F 13.2
18:00 242.9 F 11.7
20:00 255.6 F 18.2
22:00 266.7 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.9 F 29.0
2:00 272.0 F 28.3
3:00 271.7 F 28.4
4:00 272.9 F 29.0

Min 271.5 F 28.0
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 212.6 F 0.0
10:00 209.5 F 0.0
15:00 212.6 F 0.0
20:00 214.3 F 0.0
25:00 223.5 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 23:35
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:01:57
CYCLE COMPLETE

10/29/2021 09:41 AM
ACCEPTED: X
REJECTED:
Operator: U02