
Midmark M11 - v1.0.5
Total Cycles: 2905

Name: RITTER2
Sterilizer ID: V2313228
09/19/2023 12:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.6 F 0.0
2:00 137.0 F 0.0
4:00 163.8 F 1.1
6:00 185.3 F 4.9
8:00 206.5 F 9.1
10:00 223.0 F 13.4
12:00 229.3 F 7.9
14:00 240.6 F 12.6
16:00 238.0 F 10.1
18:00 237.4 F 9.2
20:00 249.0 F 15.1
22:00 260.4 F 21.4
24:00 270.8 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.7
1:00 272.5 F 29.3
2:00 271.0 F 28.6
3:00 273.0 F 29.8
4:00 272.0 F 29.7

Min 270.9 F 28.6
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.1 F 0.0
5:00 230.0 F 0.0
10:00 224.0 F 0.0
15:00 213.4 F 0.0
20:00 207.9 F 0.0
25:00 196.2 F 0.0
30:00 183.9 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 24:47
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:03:21
CYCLE COMPLETE

09/19/2023 02:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____