
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004020
15:18 14/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:14
STERILIZING 4:24
136.5 C 224kPa 4:24
136.3 C 225kPa 4:54
136.2 C 223kPa 5:24
136.4 C 224kPa 5:54
136.4 C 224kPa 6:24
136.2 C 223kPa 6:54
136.4 C 225kPa 7:24
136.4 C 225kPa 7:54
136.4 C 225kPa 7:54
VENTING 7:54
AIR DRYING 8:41
CYCLE COMPLETE 25:06

Digital Signature #
FB4E602D5AAA50A3
