

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3503

Name: RITTER2  
Sterilizer ID: V2313228  
04/08/2024 10:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 67.5 F 0.0  
2:00 68.0 F 0.0  
4:00 77.6 F 0.0  
6:00 111.1 F 0.1  
8:00 152.2 F 0.4  
10:00 170.8 F 3.4  
12:00 189.2 F 6.6  
14:00 202.2 F 10.1  
16:00 216.8 F 11.4  
18:00 225.5 F 7.4  
20:00 232.9 F 11.6  
22:00 232.9 F 8.3  
24:00 243.5 F 12.4  
26:00 243.8 F 12.5  
28:00 255.5 F 18.6  
30:00 266.0 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.8  
1:00 272.4 F 29.2  
2:00 271.5 F 29.1  
3:00 272.1 F 29.0  
4:00 271.5 F 29.1  
  
Min 271.2 F 28.6  
Max 272.5 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.8 F 0.0  
5:00 213.9 F 0.0  
10:00 207.0 F 0.0  
15:00 199.9 F 0.0  
20:00 198.5 F 0.0  
25:00 200.8 F 0.0  
30:00 202.5 F 0.0  
  
Drying Complete

FILLING: 0:36  
HEATING: 31:45  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:09:43  
CYCLE COMPLETE

04/08/2024 11:18 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_