

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1378

Name: RITTER2  
Sterilizer ID: V2313228  
08/08/2022 03:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.8 F 0.0  
2:00 187.3 F 0.2  
4:00 203.0 F 4.5  
6:00 221.8 F 10.0  
8:00 231.4 F 7.0  
10:00 240.2 F 13.0  
12:00 241.6 F 11.9  
14:00 242.3 F 11.7  
16:00 256.0 F 19.1  
18:00 268.3 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.5 F 29.4  
2:00 271.7 F 29.1  
3:00 272.5 F 29.5  
4:00 271.6 F 28.8  
  
Min 271.5 F 28.8  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.3 F 0.0  
5:00 217.2 F 0.0  
10:00 233.2 F 0.0  
15:00 231.0 F 0.0  
20:00 226.2 F 0.0  
25:00 221.9 F 0.0  
30:00 220.5 F 0.0  
  
Drying Complete

FILLING: 1:21  
HEATING: 19:09  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:57:45  
CYCLE COMPLETE

08/08/2022 04:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_