
Midmark M11 - v1.0.5
Total Cycles: 3813

Name: RITTER2
Sterilizer ID: V2313228
07/25/2024 09:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.0 F 0.0
2:00 73.1 F 0.0
4:00 78.2 F 0.0
6:00 93.9 F 0.1
8:00 131.8 F 0.2
10:00 158.0 F 1.8
12:00 182.3 F 4.6
14:00 202.1 F 7.6
16:00 213.5 F 10.8
18:00 221.2 F 7.8
20:00 227.9 F 9.1
22:00 237.8 F 13.5
24:00 238.0 F 10.9
26:00 235.4 F 9.0
28:00 248.8 F 15.2
30:00 259.3 F 21.1
32:00 269.1 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 271.5 F 29.5
2:00 272.4 F 29.6
3:00 271.3 F 29.0
4:00 271.7 F 29.7

Min 271.1 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.0 F 0.0
5:00 216.3 F 0.0
10:00 215.3 F 0.0
15:00 207.0 F 0.0
20:00 212.7 F 0.0
25:00 212.6 F 0.0
30:00 215.8 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 33:03
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:11:36
CYCLE COMPLETE

07/25/2024 10:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____