
Midmark M11 - v1.0.5
Total Cycles: 3486

Name: RITTER2
Sterilizer ID: V2313228
03/28/2024 09:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.5 F 0.0
2:00 71.1 F 0.0
4:00 78.6 F 0.0
6:00 108.0 F 0.1
8:00 142.3 F 0.4
10:00 170.3 F 3.0
12:00 189.9 F 6.1
14:00 200.7 F 10.1
16:00 216.7 F 8.4
18:00 224.6 F 7.9
20:00 235.8 F 12.4
22:00 235.5 F 9.4
24:00 239.0 F 10.0
26:00 246.3 F 13.7
28:00 257.7 F 19.9
30:00 268.2 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.6 F 29.4
2:00 271.4 F 28.8
3:00 272.8 F 29.4
4:00 271.6 F 28.7

Min 271.4 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.0 F 0.0
5:00 204.2 F 0.0
10:00 195.7 F 0.0
15:00 191.2 F 0.0
20:00 189.6 F 0.0
25:00 190.3 F 0.0
30:00 191.7 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:19
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:09:25
CYCLE COMPLETE

03/28/2024 10:40 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____