

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1717

Name: RITTER 1  
Sterilizer ID: V2178198  
03/15/2022 09:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.0 F 0.0  
2:00 74.8 F 0.0  
4:00 85.4 F 0.0  
6:00 114.2 F 0.9  
8:00 147.0 F 2.4  
10:00 173.9 F 4.4  
12:00 193.2 F 7.2  
14:00 206.1 F 10.8  
16:00 220.5 F 5.2  
18:00 226.9 F 8.2  
20:00 238.3 F 12.7  
22:00 236.7 F 9.7  
24:00 238.0 F 8.8  
26:00 248.3 F 14.2  
28:00 260.0 F 20.6  
30:00 270.4 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.1  
1:00 272.3 F 28.8  
2:00 271.5 F 28.1  
3:00 271.8 F 28.6  
4:00 272.3 F 28.6

Min 271.0 F 27.9  
Max 272.4 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.4 F 0.0  
5:00 216.9 F 0.0  
10:00 220.4 F 0.0  
15:00 227.9 F 0.0  
20:00 232.7 F 0.0  
25:00 229.9 F 0.0  
30:00 227.2 F 0.0  
Drying Complete

FILLING: 0:45  
HEATING: 30:53  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:08:40  
CYCLE COMPLETE

03/15/2022 11:05 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_