

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.1uS / 0.7ppm  
CYCLE NUMBER 001465  
13:15 09/30/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:28  
PRESSURIZING 6:09  
STERILIZING 7:41  
136.0 C 220kPa 7:41  
136.3 C 223kPa 8:11  
135.9 C 221kPa 8:41  
136.1 C 221kPa 9:11  
136.4 C 223kPa 9:41  
136.3 C 222kPa 10:11  
136.3 C 223kPa 10:41  
136.1 C 221kPa 11:11  
VENTING 11:11  
AIR DRYING 12:18  
CYCLE COMPLETE 23:05

-----  
Digital Signature #  
37359587BBCE1BD0  
-----