
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000783
9:45 29/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:31
STERILIZING 3:35
136.5 C 233kPa 3:35
136.4 C 233kPa 4:05
136.4 C 233kPa 4:35
136.3 C 233kPa 5:05
136.4 C 233kPa 5:35
136.4 C 233kPa 6:05
136.4 C 233kPa 6:35
136.4 C 233kPa 7:05
136.4 C 233kPa 7:05
VENTING 7:05
AIR DRYING 7:57
CYCLE COMPLETE 7:59

Digital Signature #
F57565485E017288
