

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000446  
8:16 01/02/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:26  
PRESSURIZING 3:31  
STERILIZING 4:45  
136.0 C 229kPa 4:45  
136.2 C 231kPa 5:15  
136.4 C 233kPa 5:45  
136.2 C 230kPa 6:15  
136.2 C 231kPa 6:45  
136.3 C 231kPa 7:15  
136.1 C 232kPa 7:45  
136.2 C 231kPa 8:15  
136.2 C 231kPa 8:15  
VENTING 8:15  
AIR DRYING 9:10  
CYCLE COMPLETE 9:26  
-----

Digital Signature #  
6025A1BAF8D149F8  
-----