
Midmark M11 - v1.0.5
Total Cycles: 437

Name: RITTER2
Sterilizer ID: V2313228
10/12/2021 09:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.8 F 0.0
2:00 148.0 F 1.6
4:00 147.5 F 4.7
6:00 172.7 F 9.2
8:00 201.2 F 6.6
10:00 220.9 F 8.9
12:00 233.8 F 10.5
14:00 238.0 F 12.3
16:00 238.7 F 11.0
18:00 251.8 F 17.5
20:00 263.1 F 24.1
22:00 272.2 F 30.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 30.4
1:00 272.1 F 30.6
2:00 272.2 F 30.7
3:00 272.3 F 30.8
4:00 272.2 F 30.7

Min 271.9 F 30.3
Max 272.7 F 31.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.8 F 0.0
5:00 227.2 F 0.0
10:00 244.0 F 0.0
15:00 232.1 F 0.0
20:00 190.2 F 0.0
25:00 176.8 F 0.0
30:00 172.6 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 22:53
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:01:05
CYCLE COMPLETE

10/12/2021 10:27 AM
ACCEPTED: X
REJECTED:
Operator: U01