

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001553  
16:40 30/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:15  
PRESSURIZING 2:24  
STERILIZING 3:20  
136.0 C 220kPa 3:20  
136.3 C 224kPa 3:50  
136.4 C 223kPa 4:20  
136.1 C 223kPa 4:50  
136.2 C 225kPa 5:20  
136.4 C 226kPa 5:50  
136.3 C 225kPa 6:20  
136.3 C 224kPa 6:50  
136.3 C 224kPa 6:50  
VENTING 6:50  
AIR DRYING 7:42  
CYCLE COMPLETE 37:42  
-----

Digital Signature #  
8D4D7E364F4EF4D0  
-----