

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2961

Name: RITTER2  
Sterilizer ID: V2313228  
10/02/2023 04:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 152.3 F 0.0  
2:00 154.8 F 0.2  
4:00 184.6 F 4.1  
6:00 203.2 F 9.2  
8:00 215.2 F 7.8  
10:00 229.5 F 9.4  
12:00 235.6 F 9.6  
14:00 244.3 F 13.2  
16:00 244.6 F 12.6  
18:00 257.7 F 19.7  
20:00 269.5 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 271.9 F 28.6  
2:00 272.4 F 28.9  
3:00 272.6 F 29.3  
4:00 271.6 F 28.5  
  
Min 271.6 F 28.4  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.2 F 0.0  
5:00 209.6 F 0.0  
10:00 218.3 F 0.0  
15:00 221.8 F 0.0  
20:00 220.1 F 0.0  
25:00 218.0 F 0.0  
30:00 216.2 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 20:54  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:59:11  
CYCLE COMPLETE

10/02/2023 05:12 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_