

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3552

Name: RITTER2  
Sterilizer ID: V2313228  
04/24/2024 09:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 138.2 F 0.0  
2:00 141.6 F 1.6  
4:00 150.9 F 4.1  
6:00 182.4 F 8.0  
8:00 203.8 F 12.5  
10:00 215.6 F 6.7  
12:00 228.5 F 11.2  
14:00 230.3 F 7.6  
16:00 241.0 F 13.1  
18:00 240.9 F 11.5  
20:00 252.3 F 17.5  
22:00 263.0 F 23.8  
24:00 271.8 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.8  
1:00 271.4 F 30.0  
2:00 271.6 F 30.1  
3:00 271.6 F 30.2  
4:00 271.3 F 30.1

Min 271.2 F 29.7  
Max 272.9 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 164.2 F 0.0  
5:00 253.2 F 0.0  
10:00 191.4 F 0.0  
15:00 189.1 F 0.0  
20:00 186.8 F 0.0  
25:00 186.6 F 0.0  
30:00 191.8 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 24:48  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 01:03:18  
CYCLE COMPLETE

04/24/2024 11:01 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_