

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001973  
11:12 22/08/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:28  
PRESSURIZING 2:37  
STERILIZING 3:33  
136.0 C 220kPa 3:33  
136.4 C 225kPa 4:03  
136.4 C 223kPa 4:33  
136.1 C 223kPa 5:03  
136.3 C 224kPa 5:33  
136.4 C 225kPa 6:03  
136.4 C 225kPa 6:33  
136.3 C 222kPa 7:03  
136.3 C 222kPa 7:03  
VENTING 7:03  
AIR DRYING 7:53  
CYCLE COMPLETE 37:53  
-----

Digital Signature #  
392AB66E7E07F84C  
-----