
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
2.0uS / 1.2ppm
CYCLE NUMBER 003730
7:59 31/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:06
STERILIZING 4:08
136.0 C 220kPa 4:08
136.0 C 223kPa 4:38
135.9 C 222kPa 5:08
136.4 C 224kPa 5:38
136.4 C 222kPa 6:08
136.3 C 221kPa 6:38
136.2 C 224kPa 7:08
136.4 C 223kPa 7:38
136.4 C 223kPa 7:38
VENTING 7:38
AIR DRYING 8:26
CYCLE COMPLETE 8:29

Digital Signature #
3075429B5E655158
