
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007718
8:18 09/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:38
STERILIZING 4:12
136.4 C 220kPa 4:12
135.7 C 214kPa 4:42
136.3 C 208kPa 5:12
136.3 C 216kPa 5:42
136.3 C 214kPa 6:12
136.0 C 217kPa 6:42
136.0 C 214kPa 7:12
136.4 C 220kPa 7:42
136.4 C 220kPa 7:42
VENTING 7:42
AIR DRYING 8:34
CYCLE COMPLETE 38:34

Digital Signature #
BC6729017F18937D
